


Technical Specification

Specification of Endomotor (Dental)

S.N.	Purchaser's Specification (Tamakoshi Hospital)	Bidders' Compliance Sheet		
		Yes/No	Pg. No. Catalogue	Remarks
	Brand			
	Type/Mode			
	Country of Origin			
1	Low noise and small vibration motor			
2	Viable colorful OLED screen			
3	Endodontic 1:1 contra angle head, 320° adjustable handpiece			
4	Rotational speed range: 120-1000rpm, torque range:0.6-4.0N.cm			
5	Editable torque, speed and FWD/REV rotation angle (30-250°)			
6	Memory Programs: 10 (M0-M9)			
7	Humanized and switchable mode of left and right-hand			
8	Power Supply: 3.7V/2000mAh ultra-large battery			
9	Charger power supply: 100-240 V, AC			
10	User Training			
10.1	Must provide user training (including how to use and maintain the equipment).			
11	Warranty			
11.1	Comprehensive warranty for 2 year.			
12	Maintenance Service During Warranty Period			
12.1	During the warranty period, supplier must ensure planned preventive maintenance (PPM) and corrective/breakdown maintenance whenever required.			
13	Installation and Commissioning			
13.1	The bidder must arrange for the equipment to be installed and commissioned by certified or qualified personnel; any prerequisites for installation to be communicated to the purchaser in advance, in detail.			


 Prof. Dr. Subha Karmacharya
 MBBS MCh MCM
 NMC No. 3574



2-113

14	Documentation			
14.1	User (operating) manual in English			
14.2	Service (Technical/Maintenance) manual in English			
14.3	List of important spare parts and accessories with their part number and costing			
14.4	Certificate of calibration and inspection from factory			
<p>The bidder must completely fill out the Technical Specification Form (TSF). Only Yes/No/All complies should not be written. The page number in the catalog of all the required parameters must be mentioned and highlighted. Failure to do so may lead to the rejection of the bid from the technical committee.</p>				

[Handwritten signature]

Prof. Dr. Suman Karmacharya
 MBBS MHCM
 NMC No. 3574



2-113